

Quality Certificate

PTFE Syringe Filter,0.22µm, 30mm,Non-Sterile	Part Number:LC-OL-SF-PTFE030022NB
	Lot Number:170707037

ISO 9001 Quality Standard



The product was manufactured in the facility whose Quality Management System is approved by an accredited registering body to the ISO 9001 Quality Systems Standard.

Syringe Filter Lot Characterization

Membrane material	PTFE	Filtration area(cm ²)	4.9
Pore size	0.22µm	Holdup volume (µl)	<120
Wettability	Hydrophobic	Volume Throughput (ml)	<100
Housing Material	Polypropylene	Maximum Operating Temperature(°C)	130
Inlet/Outlet	Female Luer Lock inlet and Male Luer slip outlet	Bubble point(psi)	20.3-23.2
Burst pressure(psi)	87	Flow rate(ml/min@10psi)	70-98

Bubble point and Flow rate were tested by ethyl alcohol.

HPLC Extractable Test (70% Acetonitrile,254nm)



Specification	Result
No detectable peaks	Pass

We hereby certify that the above stated filters are in compliance with all applicable instructions. The document contains information that may be confidential and is intended only for the use of the addressee.

Tested By:  2017-07-14
QC Technician